

Technical Specifications For Laboratory Equipments

Item Description: Heamatocrit Centrifuge

Class No. : (18-270)

Quantity : ()

No.	Technical Specifications	Min Requirement	Bidder Offer
1.	Manufacturer	Please specify manufacturer and country of origin	
2.	Model Number	Please specify model number of the offered equipment	
3.	Safety standard	FDA approval or CE marking	
4.	Type	Bench top	
5.	Speed , rpm	Not less than 13,000rpm (with rotor)	
6.	Capacity	24 capillary tubes (75mm capillary)	
7.	Control	Microprocessor	
8.	Display	Digital	
9.	Timer	Digital	
10.	Braking type	Electric type	
11.	Braking time, sec	Less than 30sec.	
12.	Lid interlock	Included	
13.	Alert indicators	Lid open, power, start/stop ...etc.	
14.	Noise level, dB	Less than 70dB	
15.	Reader	Included	
16.	Power supply	220 V / 50 Hz	

Validly of offer:

Warrant:

Delivery: